

PN: 402 14712

Rev. B

Date of last revision:

2009/07/10

4.0" width

3.0" length

Languages:

English

German

Spanish

Italian

French

Portuguese

Danish: Forventede værdier

Greek

Swedish: Förväntade värden

Front

Cholestech REF 14-713
LDX LOT C17 7959B

**hsCRP Calibration
Verification Controls**

EXPECTED VALUES / Erwartete Werte / Valores esperados / Valori previsti / Valeurs attendues / Valores esperados / Forventede værdier / Αναμενόμενες τιμές // Förväntade värden

hsCRP Calibration Verification Controls LOT 0905-105	\bar{x} (Mean)	$3 \times \sigma$ (3 Stdev)
L1 0905016	0.56	0.30-1.05
L2 0905017	2.81	1.40-5.09
L3 0905018	4.76	3.00-7.79

Back

EC REP

AR-MED
A Division of Quintiles Ltd
Runnymede Malthouse Business
Centre off Hummer Road
Egham, Surrey TW20 9BD
United Kingdom


Cholestech Corporation
Hayward, CA 94545 U.S.A

402 14712 Rev. B