

Cholestech REF 14-713
LDX[®] LOT C17 7656

hsCRP Calibration Verification Controls

EXPECTED VALUES / Erwartete Werte / Valores
esperados / Valori previsti / Valeurs attendues /
Valores esperados / Forventede værdier /
Αναμενόμενες τιμές // Förväntade värden

hsCRP Calibration Verification Controls LOT 0905105	\bar{x} (Mean)	$3 \times \sigma$ (3 Stdev)
L1 0905016	0.77	0.35 - 1.43
L2 0905017	3.70	2.41 - 5.93
L3 0905018	5.78	3.66 - 8.00

U.S.A.:
Manufactured for
Cholestech Corporation,
3347 Investment Blvd.,
Hayward, CA 9545 U.S.A

EC REP

AR-MED
A Division of Quintiles Ltd
Runnymede Malthouse Business
Centre off Hummer Road
Egham, Surry TW20 9BD
United Kingdom



Cholestech Corporation
Hayward, CA 9545 U.S.A