

**Cholestech** REF 14-713  
**LDX**® LOT C17 7794

## hsCRP Calibration Verification Controls

EXPECTED VALUES / Erwartete Werte / Valores  
esperados / Valori previsti / Valeurs attendues /  
Valores esperados / Forventede værdier /  
Αναμενόμενες τιμές // Förväntade värden

hsCRP Calibration Verification Controls LOT 0905105	$\bar{x}$ (Mean)	$3 \times \sigma$ (3 Stdev)
L1 0905016	0.57	0.20-1.15
L2 0905017	2.87	1.45-5.18
L3 0905019	4.71	2.30-8.00

**EC REP**

AR-MED  
A Division of Quintiles Ltd  
Runnymede Malthouse Business  
Centre off Hummer Road  
Egham, Surrey TW20 9BD  
United Kingdom



Cholestech Corporation  
Hayward, CA 94545 U.S.A