

PN: 402 14712

Rev. B

Date of last revision:
2009/07/10

4.0" width
3.0" length

Languages:

English
German
Spanish
Italian
French
Portuguese
Danish: Forventede værdier
Greek
Swedish: Förväntade värden

Front

Cholestech REF 14-713
LDX LOT C17 7841

**hsCRP Calibration
Verification Controls**


EXPECTED VALUES / Erwartete Werte / Valores esperados / Valori previsti / Valeurs attendues / Valores esperados / Forventede værdier / Αναμενόμενες τιμές // Förväntade värden

hsCRP Calibration Verification Controls LOT 0905-105	\bar{x} (Mean)	3 x σ (3 Stdev)
L1 0905016	0.64	0.23-1.27
L2 0905017	3.18	1.53-5.84
L3 0905018	5.05	2.39-8.00

Back

EC REP

AR-MED
A Division of Quintiles Ltd
Runnymede Malthouse Business
Centre off Hummer Road
Egham, Surrey TW20 9BD
United Kingdom



Cholestech Corporation
Hayward, CA 94545 U.S.A

402 14712 Rev. B