

**CHOLESTECH LDX®
CALIBRATION VERIFICATION MATERIAL— RESULT SHEET**

Lot No.: 9287

Issue Date: 2009-09-25

Expiration Date: 2010-10-14

Institution: _____

Contact Person: _____

Date Analyzed: _____

Address: _____

Cholestech LDX s/n: _____

City/State/Zip: _____

Cassette Lot No.: _____

Telephone: _____

Operator: _____

Fax: _____

TRIGLYCERIDE RESULTS (mg/dL)

Sample ID	Results		Average	Target Mg/dL	Range	
	#1	#2			Low	High
Level 1				143.8	116	171
Level 2				181.5	147	216
Level 3				271.7	220	324
Level 4				412.1	333	491

TOTAL CHOLESTEROL RESULTS (mg/dL)

Sample ID	Results		Average	Target Mg/dL	Range	
	#1	#2			Low	High
Level 1				141.9	118	166
Level 2				214.4	178	251
Level 3				324.8	270	380
Level 4				384.1	319	449

GLUCOSE RESULTS (mg/dL)

Sample ID	Results		Average	Target Mg/dL	Range	
	#1	#2			Low	High
Level 1				88.8	72	106
Level 2				205.8	166	245
Level 3				316.6	256	377
Level 4				427.2	345	500

HDL CHOLESTEROL RESULTS (mg/dL)

Sample ID	Results		Average	Target Mg/dL	Range	
	#1	#2			Low	High
Level 1				24.8	19	30
Level 2				36.7	29	45
Level 3				52.2	41	64
Level 4				73.6	57	90

Streck Part Number: 350514

FAX results to Cholestech Technical Service (510) 732-7227 or mail to:
Cholestech Corporation, 3347 Investment Blvd., Hayward, CA 94545-3808
Attn: Cholestech Technical Service

PROPRIETARY INFORMATION:

Data herein is not to be disclosed without permission of Cholestech Corporation.